

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
 Data File : VX017596.D
 Acq On : 25 Jul 2020 15:15
 Operator : JC/SP
 Sample : L3382-23
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 70 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CTR-024

Quant Time: Jul 27 03:50:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	411526	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	702761	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	729316	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	381362	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	285381	53.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.98%	
35) Dibromofluoromethane	5.47	113	234666	49.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.86%	
50) Toluene-d8	8.70	98	890277	52.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.02%	
62) 4-Bromofluorobenzene	11.12	95	385908	56.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.74%	
Target Compounds						
16) Acetone	2.42	43	49926	21.458	ug/l	94
18) Methyl Acetate	2.75	43	7983	1.457	ug/l #	82
20) Methylene Chloride	2.84	84	40334	8.129	ug/l	99
43) Isopropyl Acetate	6.42	43	139906	11.498	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
Data File : VX017596.D
Acq On : 25 Jul 2020 15:15
Operator : JC/SP
Sample : L3382-23
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 70 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CTR-024

Quant Time: Jul 27 03:50:23 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

