

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
 Data File : VX017597.D
 Acq On : 25 Jul 2020 15:38
 Operator : JC/SP
 Sample : L3383-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 71 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CRT-013

Quant Time: Jul 27 03:53:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	421701	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	689072	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	727867	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	413987	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	294424	54.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.70%	
35) Dibromofluoromethane	5.46	113	243387	52.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.62%	
50) Toluene-d8	8.70	98	858491	51.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.28%	
62) 4-Bromofluorobenzene	11.12	95	385239	57.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.80%	
Target Compounds						
16) Acetone	2.42	43	34679	14.545	ug/l	93
18) Methyl Acetate	2.75	43	8491	1.512	ug/l	95
20) Methylene Chloride	2.84	84	40146	7.858	ug/l	98
43) Isopropyl Acetate	6.42	43	136607	11.450	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
Data File : VX017597.D
Acq On : 25 Jul 2020 15:38
Operator : JC/SP
Sample : L3383-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 71 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CRT-013

Quant Time: Jul 27 03:53:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

