

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
 Data File : VX017599.D
 Acq On : 25 Jul 2020 16:23
 Operator : JC/SP
 Sample : L3383-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 73 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CRT-009

Quant Time: Jul 27 03:57:50 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	414234	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	737433	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	768094	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	402943	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	315343	59.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.52%	
35) Dibromofluoromethane	5.46	113	246926	50.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.14%	
50) Toluene-d8	8.70	98	901454	50.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.34%	
62) 4-Bromofluorobenzene	11.12	95	390997	54.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.82%	
Target Compounds						
16) Acetone	2.42	43	31116	13.286	ug/l #	87
18) Methyl Acetate	2.76	43	8058	1.461	ug/l #	88
20) Methylene Chloride	2.84	84	40168	8.029	ug/l	95
43) Isopropyl Acetate	6.42	43	149306	11.694	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
Data File : VX017599.D
Acq On : 25 Jul 2020 16:23
Operator : JC/SP
Sample : L3383-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 73 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CRT-009

Quant Time: Jul 27 03:57:50 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

