

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
 Data File : VX017600.D
 Acq On : 25 Jul 2020 16:46
 Operator : JC/SP
 Sample : L3383-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 74 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CRT-022

Quant Time: Jul 27 04:00:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	423726	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	655152	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	728692	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	378407	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	293976	54.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.02%	
35) Dibromofluoromethane	5.46	113	252848	57.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.42%	
50) Toluene-d8	8.70	98	832161	52.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.30%	
62) 4-Bromofluorobenzene	11.12	95	361838	57.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.40%	
Target Compounds						
16) Acetone	2.42	43	31614	13.197	ug/l	90
18) Methyl Acetate	2.75	43	8227	1.458	ug/l #	70
20) Methylene Chloride	2.84	84	39908	7.760	ug/l	95
43) Isopropyl Acetate	6.42	43	137528	12.124	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
Data File : VX017600.D
Acq On : 25 Jul 2020 16:46
Operator : JC/SP
Sample : L3383-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 74 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CRT-022

Quant Time: Jul 27 04:00:09 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

