

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
 Data File : VX017603.D
 Acq On : 25 Jul 2020 17:54
 Operator : JC/SP
 Sample : L3383-27
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 77 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CRT-014

Quant Time: Jul 27 04:07:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 24 05:48:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	426201	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	670996	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	779222	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	371620	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	300509	54.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.78%	
35) Dibromofluoromethane	5.47	113	248897	55.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.92%	
50) Toluene-d8	8.70	98	832810	51.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.88%	
62) 4-Bromofluorobenzene	11.12	95	358813	55.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.76%	
Target Compounds						
16) Acetone	2.42	43	35074	14.556	ug/l	94
18) Methyl Acetate	2.75	43	8383	1.477	ug/l	92
20) Methylene Chloride	2.84	84	42869	8.377	ug/l	99
43) Isopropyl Acetate	6.42	43	139456	12.004	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX072420\
Data File : VX017603.D
Acq On : 25 Jul 2020 17:54
Operator : JC/SP
Sample : L3383-27
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 77 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CRT-014

Quant Time: Jul 27 04:07:38 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072320W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 24 05:48:09 2020
Response via : Initial Calibration

