

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072423\
 Data File : VX036622.D
 Acq On : 24 Jul 2023 11:18
 Operator : JC/MD
 Sample : VX0724WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0724WBL01

Quant Time: Jul 25 06:33:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	181664	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	322227	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	301661	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	138024	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	122724	45.656	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.320%
35) Dibromofluoromethane	5.385	113	103282	46.632	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.260%
50) Toluene-d8	8.647	98	386821	49.250	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.500%
62) 4-Bromofluorobenzene	11.079	95	151455	51.111	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.220%
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.788	84	2488	1.003	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072423\
Data File : VX036622.D
Acq On : 24 Jul 2023 11:18
Operator : JC/MD
Sample : VX0724WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0724WBL01

Quant Time: Jul 25 06:33:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

