



BFB

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072423\
Data File : VX036619.D
Acq On : 24 Jul 2023 09:20
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Title : SW846 8260
Last Update : Fri Jul 21 12:14:12 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	19.0	11757	PASS
75	95	30	60	51.7	31968	PASS
95	95	100	100	100.0	61811	PASS
96	95	5	9	6.9	4260	PASS
173	174	0.00	2	0.8	336	PASS
174	95	50	100	71.6	44272	PASS
175	174	5	9	7.3	3210	PASS
176	174	95	101	97.2	43053	PASS
177	176	5	9	6.4	2743	PASS