

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072424\
 Data File : VX042473.D
 Acq On : 23 Jul 2024 15:44
 Operator : JC/MD
 Sample : P3264-11ME
 Misc : 5.18g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 A4CL4ME

Quant Time: Jul 24 03:48:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML070324WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jul 24 03:44:32 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	224723	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	206003	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	99864	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	39835	30.454	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	60.900%
7) Chloroethane-d5	1.648	69	23482	28.512	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	57.020%#
11) 1,1-Dichloroethene-d2	2.282	65	16884	29.493	ug/L	-0.01
Spiked Amount	50.000	Range	60 - 125	Recovery	=	58.980%#
21) 2-Butanone-d5	4.489	46	80914	85.840	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	85.840%
24) Chloroform-d	5.056	84	106232	40.393	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.780%
26) 1,2-Dichloroethane-d4	5.952	65	66846	43.062	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.120%
32) Benzene-d6	5.964	84	226406	42.520	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.040%
36) 1,2-Dichloropropane-d6	7.306	67	69195	43.181	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	86.360%
41) Toluene-d8	8.647	98	196921	39.685	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	79.380%#
43) trans-1,3-Dichloroprop...	8.952	79	25287	36.513	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.020%
47) 2-Hexanone-d5	9.391	63	65326	89.811	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	89.810%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	100477	42.413	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	84.820%
66) 1,2-Dichlorobenzene-d4	12.323	152	85079	43.862	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.720%
Target Compounds						
					Qvalue	
15) Methyl Acetate	2.721	43	15680	9.597	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072424\
Data File : VX042473.D
Acq On : 23 Jul 2024 15:44
Operator : JC/MD
Sample : P3264-11ME
Misc : 5.18g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
A4CL4ME

Quant Time: Jul 24 03:48:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML070324WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jul 24 03:44:32 2024
Response via : Initial Calibration

