




Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072424\
Data File : VX042470.D
Acq On : 23 Jul 2024 14:34
Operator : JC/MD
Sample : P3263-19ME
Misc : 5.15g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
A4CK9ME

Manual IntegrationsAPPROVED

Reviewed By :John Carlone 07/24/2024
Supervised By :Mahesh Dadoda 07/24/2024

Quant Time: Jul 24 03:48:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML070324WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jul 24 03:44:32 2024
Response via : Initial Calibration



TIC: VX042470.D\data.ms

(4) Vinyl Chloride-d3 (S)

1.392min (+ 0.024) 21.01 ug/L

response 29155

Ion	Exp%	Act%
65.00	100.00	100.00
67.00	31.70	40.84
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072424\
Data File : VX042470.D
Acq On : 23 Jul 2024 14:34
Operator : JC/MD
Sample : P3263-19ME
Misc : 5.15g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
A4CK9ME

Manual IntegrationsAPPROVED

Reviewed By :John Carlone 07/24/2024
Supervised By :Mahesh Dadoda 07/24/2024

Quant Time: Jul 24 03:48:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML070324WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jul 24 03:44:32 2024
Response via : Initial Calibration



TIC: VX042470.D\data.ms

(4) Vinyl Chloride-d3 (S)

1.392min (+ 0.024) 26.05 ug/L m

response 36156

Ion	Exp%	Act%
65.00	100.00	100.00
67.00	31.70	32.93
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072424\
Data File : VX042470.D
Acq On : 23 Jul 2024 14:34
Operator : JC/MD
Sample : P3263-19ME
Misc : 5.15g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
A4CK9ME

Manual IntegrationsAPPROVED

Reviewed By :John Carlone 07/24/2024
Supervised By :Mahesh Dadoda 07/24/2024

Quant Time: Jul 24 03:48:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML070324WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jul 24 03:44:32 2024
Response via : Initial Calibration



TIC: VX042470.D\data.ms

(4) Vinyl Chloride-d3 (S)

1.392min (+ 0.024) 26.04 ug/L m

response 36130

Ion	Exp%	Act%
65.00	100.00	100.00
67.00	31.70	32.95
0.00	0.00	0.00
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072424\
Data File : VX042470.D
Acq On : 23 Jul 2024 14:34
Operator : JC/MD
Sample : P3263-19ME
Misc : 5.15g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
A4CK9ME

Manual IntegrationsAPPROVED

Quant Time: Jul 24 03:48:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML070324WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jul 24 03:44:32 2024
Response via : Initial Calibration

Reviewed By :John Carlone 07/24/2024
Supervised By :Mahesh Dadoda 07/24/2024



TIC: VX042470.D\data.ms

(7) Chloroethane-d5 (S)

1.648min (+ 0.000) 6.17 ug/L

response 5392

Ion	Exp%	Act%
69.00	100.00	100.00
71.00	33.20	40.97
51.00	12.70	11.44
0.00	0.00	0.00




Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072424\
Data File : VX042470.D
Acq On : 23 Jul 2024 14:34
Operator : JC/MD
Sample : P3263-19ME
Misc : 5.15g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
A4CK9ME

Manual IntegrationsAPPROVED

Reviewed By :John Carlone 07/24/2024
Supervised By :Mahesh Dadoda 07/24/2024

Quant Time: Jul 24 03:48:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML070324WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jul 24 03:44:32 2024
Response via : Initial Calibration



TIC: VX042470.D\data.ms

(7) Chloroethane-d5 (S)

1.697min (+ 0.049) 27.07 ug/L m

response 23656

Ion	Exp%	Act%
69.00	100.00	100.00
71.00	33.20	9.34#
51.00	12.70	2.61#
0.00	0.00	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072424\
 Data File : VX042470.D
 Acq On : 23 Jul 2024 14:34
 Operator : JC/MD
 Sample : P3263-19ME
 Misc : 5.15g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 A4CK9ME

Manual IntegrationsAPPROVED

Quant Time: Jul 24 03:48:04 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML070324WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Jul 24 03:44:32 2024
 Response via : Initial Calibration

Reviewed By :John Carlone 07/24/2024
 Supervised By :Mahesh Dadoda 07/24/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	238410	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	215490	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	107323	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.392	65	36156m	26.055	ug/L	0.02
Spiked Amount 50.000	Range 60	- 135	Recovery	=	52.100%#	
7) Chloroethane-d5	1.697	69	23656m	27.075	ug/L	0.05
Spiked Amount 50.000	Range 70	- 130	Recovery	=	54.140%#	
11) 1,1-Dichloroethene-d2	2.270	65	15607	25.697	ug/L	-0.02
Spiked Amount 50.000	Range 60	- 125	Recovery	=	51.400%#	
21) 2-Butanone-d5	4.501	46	81221	81.219	ug/L	0.04
Spiked Amount 100.000	Range 40	- 130	Recovery	=	81.220%	
24) Chloroform-d	5.056	84	103629	37.141	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	74.280%	
26) 1,2-Dichloroethane-d4	5.952	65	67350	40.895	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	81.800%	
32) Benzene-d6	5.964	84	218187	39.172	ug/L	0.00
Spiked Amount 50.000	Range 70	- 125	Recovery	=	78.340%	
36) 1,2-Dichloropropane-d6	7.312	67	69522	41.475	ug/L	0.00
Spiked Amount 50.000	Range 70	- 120	Recovery	=	82.960%	
41) Toluene-d8	8.647	98	195220	37.611	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	75.220%#	
43) trans-1,3-Dichloroprop...	8.952	79	25273	34.886	ug/L	0.00
Spiked Amount 50.000	Range 60	- 125	Recovery	=	69.780%	
47) 2-Hexanone-d5	9.397	63	70033	92.044	ug/L	0.01
Spiked Amount 100.000	Range 45	- 130	Recovery	=	92.040%	
56) 1,1,2,2-Tetrachloroeth...	11.195	84	102857	41.506	ug/L	0.00
Spiked Amount 50.000	Range 65	- 120	Recovery	=	83.020%	
66) 1,2-Dichlorobenzene-d4	12.323	152	85479	41.006	ug/L	0.00
Spiked Amount 50.000	Range 80	- 120	Recovery	=	82.020%	
Target Compounds						
					Qvalue	
13) Acetone	2.422	43	5312	7.787	ug/L	90
15) Methyl Acetate	2.721	43	52533	30.307	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072424\
Data File : VX042470.D
Acq On : 23 Jul 2024 14:34
Operator : JC/MD
Sample : P3263-19ME
Misc : 5.15g/5.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
A4CK9ME

Manual IntegrationsAPPROVED

Quant Time: Jul 24 03:48:04 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML070324WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Jul 24 03:44:32 2024
Response via : Initial Calibration

Reviewed By :John Carlone 07/24/2024
Supervised By :Mahesh Dadoda 07/24/2024

