

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072425\
 Data File : VX047106.D
 Acq On : 24 Jul 2025 10:13
 Operator : JC/MD
 Sample : VX0724MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0724MBL01

Quant Time: Jul 25 01:18:44 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	340709	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	609396	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	574857	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	291142	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	206085	48.557	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.120%
35) Dibromofluoromethane	5.391	113	168924	44.895	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.800%
50) Toluene-d8	8.646	98	604794	49.634	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.260%
62) 4-Bromofluorobenzene	11.079	95	250976	47.795	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072425\
Data File : VX047106.D
Acq On : 24 Jul 2025 10:13
Operator : JC/MD
Sample : VX0724MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0724MBL01

Quant Time: Jul 25 01:18:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 22 03:59:51 2025
Response via : Initial Calibration

