

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072425\
 Data File : VX047111.D
 Acq On : 24 Jul 2025 12:03
 Operator : JC/MD
 Sample : Q2682-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TOWER-2

Quant Time: Jul 25 01:21:21 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	438898	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	812279	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	775011	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	390837	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	311905	57.049	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	114.100%
35) Dibromofluoromethane	5.397	113	257221	51.288	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.580%
50) Toluene-d8	8.653	98	922306	56.786	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	113.580%#
62) 4-Bromofluorobenzene	11.079	95	385676	55.102	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	110.200%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	9911	4.525	ug/l	93
30) Chloroform	5.117	83	3726	0.343	ug/l	95
47) Bromodichloromethane	7.836	83	3253	0.352	ug/l #	88
60) Dibromochloromethane	9.525	129	3453	0.516	ug/l	97
71) Bromoform	10.805	173	2564	0.552	ug/l #	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072425\
Data File : VX047111.D
Acq On : 24 Jul 2025 12:03
Operator : JC/MD
Sample : Q2682-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TOWER-2

Quant Time: Jul 25 01:21:21 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 22 03:59:51 2025
Response via : Initial Calibration

