

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072425\
 Data File : VX047120.D
 Acq On : 24 Jul 2025 15:12
 Operator : JC/MD
 Sample : Q2481-21 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CC0627-SFBL

Quant Time: Jul 25 01:24:43 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	488832	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	892867	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	845132	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	418999	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	302516	49.680	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.360%
35) Dibromofluoromethane	5.397	113	100894	18.302	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	36.600%#
50) Toluene-d8	8.653	98	856632	47.982	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.960%
62) 4-Bromofluorobenzene	11.079	95	362137	47.069	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.140%
Target Compounds						
16) Acetone	2.392	43	7716	3.163	ug/l	Qvalue # 80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072425\
Data File : VX047120.D
Acq On : 24 Jul 2025 15:12
Operator : JC/MD
Sample : Q2481-21 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CC0627-SFBL

Quant Time: Jul 25 01:24:43 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 22 03:59:51 2025
Response via : Initial Calibration

