

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072425\
 Data File : VX047124.D
 Acq On : 24 Jul 2025 16:37
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jul 25 01:26:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	356374	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	660591	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	656103	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	342547	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	272537	61.392	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	122.780%#
35) Dibromofluoromethane	5.397	113	222488	54.549	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	109.100%
50) Toluene-d8	8.653	98	813959	61.622	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	123.240%#
62) 4-Bromofluorobenzene	11.079	95	354390	62.258	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	124.520%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072425\
Data File : VX047124.D
Acq On : 24 Jul 2025 16:37
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jul 25 01:26:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 22 03:59:51 2025
Response via : Initial Calibration

