

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072425\
 Data File : VX047125.D
 Acq On : 24 Jul 2025 16:58
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jul 25 01:26:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	353076	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	661776	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	628560	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	324837	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	231880	52.721	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.440%
35) Dibromofluoromethane	5.397	113	188481	46.128	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.260%
50) Toluene-d8	8.653	98	699165	52.837	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.680%
62) 4-Bromofluorobenzene	11.079	95	295206	51.768	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072425\
Data File : VX047125.D
Acq On : 24 Jul 2025 16:58
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jul 25 01:26:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 22 03:59:51 2025
Response via : Initial Calibration

