

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072425\
 Data File : VX047112.D
 Acq On : 24 Jul 2025 12:24
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jul 25 01:21:45 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	444294	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	825514	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	779111	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	403323	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	301447	54.467	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	108.940%
35) Dibromofluoromethane	5.397	113	245823	48.229	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.460%
50) Toluene-d8	8.653	98	887521	53.768	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	107.540%
62) 4-Bromofluorobenzene	11.079	95	370685	52.111	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.220%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072425\
Data File : VX047112.D
Acq On : 24 Jul 2025 12:24
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jul 25 01:21:45 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 22 03:59:51 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 71
Delta R.T. 0.000 min
Lab File: VX047112.D
Acq: 24 Jul 2025 12:24




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion:168 Resp: 444294
Ion Ratio Lower Upper
168 100
99 65.0 48.0 72.0



#33
1,2-Dichloroethane-d4
Concen: 54.467 ug/l
RT: 5.964 min Scan# 800
Delta R.T. 0.000 min
Lab File: VX047112.D
Acq: 24 Jul 2025 12:24

Tgt Ion: 65 Resp: 301447
Ion Ratio Lower Upper
65 100
67 51.4 0.0 104.8









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.769 min Scan# 911
Delta R.T. 0.000 min
Lab File: VX047112.D
Acq: 24 Jul 2025 12:24




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion:114 Resp: 825514
Ion Ratio Lower Upper
114 100
63 21.9 0.0 43.2
88 16.5 0.0 33.8



#35
Dibromofluoromethane
Concen: 48.229 ug/l
RT: 5.397 min Scan# 707
Delta R.T. 0.000 min
Lab File: VX047112.D
Acq: 24 Jul 2025 12:24

Tgt Ion:113 Resp: 245823
Ion Ratio Lower Upper
113 100
111 103.1 82.6 124.0
192 17.4 14.9 22.3









#50
Toluene-d8
Concen: 53.768 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. 0.001 min
Lab File: VX047112.D
Acq: 24 Jul 2025 12:24




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion: 98 Resp: 887521
Ion Ratio Lower Upper
98 100
100 66.1 53.3 79.9



#62
4-Bromofluorobenzene
Concen: 52.111 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX047112.D
Acq: 24 Jul 2025 12:24

Tgt Ion: 95 Resp: 370685
Ion Ratio Lower Upper
95 100
174 72.2 0.0 144.6
176 68.9 0.0 139.6





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.055 min Scan# 1471

Delta R.T. 0.000 min

Lab File: VX047112.D

Acq: 24 Jul 2025 12:24

Instrument :

MSVOA_X

ClientSampleId :

IBLK





Tgt Ion:117 Resp: 779111

Ion Ratio Lower Upper

117 100

82 56.1 45.4 68.2

119 32.4 26.1 39.1



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 1793

Delta R.T. 0.001 min

Lab File: VX047112.D

Acq: 24 Jul 2025 12:24

Tgt Ion:152 Resp: 403323

Ion Ratio Lower Upper

152 100

115 62.4 45.6 136.7

150 157.4 0.0 354.6

