

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072518\
 Data File : VX003651.D
 Acq On : 25 Jul 2018 11:00
 Operator : JC/SP
 Sample : VX0725WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0725WBL01

Quant Time: Jul 26 01:30:58 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072418W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 25 04:33:52 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	107272	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	159130	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	151207	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	83165	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	82738	52.07	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.14%	
35) Dibromofluoromethane	5.50	113	71793	55.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.38%	
50) Toluene-d8	8.72	98	258939	52.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.32%	
62) 4-Bromofluorobenzene	11.14	95	90215	49.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.64%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072518\
Data File : VX003651.D
Acq On : 25 Jul 2018 11:00
Operator : JC/SP
Sample : VX0725WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0725WBL01

Quant Time: Jul 26 01:30:58 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072418W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 25 04:33:52 2018
Response via : Initial Calibration

