

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072518\
 Data File : VX003658.D
 Acq On : 25 Jul 2018 14:58
 Operator : JC/SP
 Sample : J4073-11DL 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE108D2-071718DL

Quant Time: Jul 26 01:49:20 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072418W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 25 04:33:52 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	91988	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	143810	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	133384	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	88213	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	66461	48.77	ug/l	0.00
Spiked Amount			Recovery	=	97.54%	
35) Dibromofluoromethane	5.51	113	53878	45.83	ug/l	0.00
Spiked Amount			Recovery	=	91.66%	
50) Toluene-d8	8.72	98	203091	45.27	ug/l	0.00
Spiked Amount			Recovery	=	90.54%	
62) 4-Bromofluorobenzene	11.14	95	69480	42.46	ug/l	0.00
Spiked Amount			Recovery	=	84.92%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
44) Trichloroethene	7.22	130	99707	74.76	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072518\
Data File : VX003658.D
Acq On : 25 Jul 2018 14:58
Operator : JC/SP
Sample : J4073-11DL 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE108D2-071718DL

Quant Time: Jul 26 01:49:20 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072418W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 25 04:33:52 2018
Response via : Initial Calibration

