

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072518\
 Data File : VX003665.D
 Acq On : 25 Jul 2018 18:37
 Operator : JC/SP
 Sample : J4073-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE126D3-071718

Quant Time: Jul 26 02:05:05 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072418W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 25 04:33:52 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	90329	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	140288	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	148350	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	73136	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	71741	53.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.24%	
35) Dibromofluoromethane	5.51	113	56654	49.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.80%	
50) Toluene-d8	8.72	98	232202	53.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.10%	
62) 4-Bromofluorobenzene	11.14	95	82201	51.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.98%	
Target Compounds						
16) Acetone	2.45	43	3398	5.41	ug/l	92
44) Trichloroethene	7.22	130	3933	3.02	ug/l	97
64) Tetrachloroethene	9.34	164	1497	1.10	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072518\
Data File : VX003665.D
Acq On : 25 Jul 2018 18:37
Operator : JC/SP
Sample : J4073-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE126D3-071718

Quant Time: Jul 26 02:05:05 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072418W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 25 04:33:52 2018
Response via : Initial Calibration

