

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072518\
 Data File : VX003666.D
 Acq On : 25 Jul 2018 19:04
 Operator : JC/SP
 Sample : J4073-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE123D2-071818

Quant Time: Jul 27 15:02:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072418W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 25 04:33:52 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	90648	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	138754	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	129697	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	83153	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	70558	52.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.10%	
35) Dibromofluoromethane	5.51	113	59993	52.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.78%	
50) Toluene-d8	8.72	98	231863	53.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.12%	
62) 4-Bromofluorobenzene	11.14	95	78170	49.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.02%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
44) Trichloroethene	7.21	130	1997	1.55	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072518\
Data File : VX003666.D
Acq On : 25 Jul 2018 19:04
Operator : JC/SP
Sample : J4073-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE123D2-071818

Quant Time: Jul 27 15:02:55 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072418W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 25 04:33:52 2018
Response via : Initial Calibration

