

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072519\
 Data File : VX011072.D
 Acq On : 25 Jul 2019 12:56
 Operator : JC/SP
 Sample : K3973-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB203-EB-20190723

Quant Time: Jul 26 01:57:20 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 24 05:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	211795	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	337089	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	312030	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	144771	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	114926	51.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.02%	
35) Dibromofluoromethane	5.50	113	107064	50.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.08%	
50) Toluene-d8	8.71	98	405479	50.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.06%	
62) 4-Bromofluorobenzene	11.13	95	141366	47.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.86%	
Target Compounds						
16) Acetone	2.44	43	10504	10.322	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072519\
Data File : VX011072.D
Acq On : 25 Jul 2019 12:56
Operator : JC/SP
Sample : K3973-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB203-EB-20190723

Quant Time: Jul 26 01:57:20 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 24 05:39:23 2019
Response via : Initial Calibration

