

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072519\
 Data File : VX011073.D
 Acq On : 25 Jul 2019 13:19
 Operator : JC/SP
 Sample : K3973-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB203-GW-198-200

Quant Time: Jul 26 02:08:14 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 24 05:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	204944	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	326961	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	306449	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	143536	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	112144	51.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.88%	
35) Dibromofluoromethane	5.50	113	102611	49.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.88%	
50) Toluene-d8	8.71	98	391702	50.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.64%	
62) 4-Bromofluorobenzene	11.13	95	140406	49.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.16%	
Target Compounds						
16) Acetone	2.44	43	5763	5.852	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072519\
Data File : VX011073.D
Acq On : 25 Jul 2019 13:19
Operator : JC/SP
Sample : K3973-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB203-GW-198-200

Quant Time: Jul 26 02:08:14 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 24 05:39:23 2019
Response via : Initial Calibration

