

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072522\
 Data File : VX030196.D
 Acq On : 25 Jul 2022 17:10
 Operator : JC/MD
 Sample : VX0725WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0725WBL01

Quant Time: Jul 25 17:58:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	149178	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	254175	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	253542	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	139453	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	155803	51.962	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	103.920%
35) Dibromofluoromethane	5.385	113	127782	52.590	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.180%
50) Toluene-d8	8.653	98	479941	51.005	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.000%
62) 4-Bromofluorobenzene	11.085	95	167417	47.941	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072522\
Data File : VX030196.D
Acq On : 25 Jul 2022 17:10
Operator : JC/MD
Sample : VX0725WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0725WBL01

Quant Time: Jul 25 17:58:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

