

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072522\
 Data File : VX030202.D
 Acq On : 25 Jul 2022 19:59
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 26 01:38:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	96	0.00
2 T	Dichlorodifluoromethane	0.820	0.881	-7.4	97	0.00
3 P	Chloromethane	0.811	0.835	-3.0	99	0.00
4 C	Vinyl Chloride	0.819	0.813	0.7#	96	0.00
5 T	Bromomethane	0.355	0.298	16.1	95	0.00
6 T	Chloroethane	0.470	0.470	0.0	99	0.00
7 T	Trichlorofluoromethane	1.079	1.095	-1.5	98	0.00
8 T	Diethyl Ether	0.422	0.426	-0.9	99	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.799	0.785	1.8	98	0.00
10 T	Methyl Iodide	0.696	0.834	-19.8	94	0.00
11 T	Tert butyl alcohol	0.245	0.260	-6.1	102	0.00
12 CM	1,1-Dichloroethene	0.781	0.780	0.1#	96	0.00
13 T	Acrolein	0.060	0.048	20.0	84	0.00
14 T	Allyl chloride	1.612	1.563	3.0	97	0.00
15 T	Acrylonitrile	0.606	0.617	-1.8	96	0.00
16 T	Acetone	0.515	0.515	0.0	98	0.00
17 T	Carbon Disulfide	2.108	1.990	5.6	93	0.00
18 T	Methyl Acetate	1.522	1.485	2.4	99	0.00
19 T	Methyl tert-butyl Ether	2.676	2.753	-2.9	96	0.00
20 T	Methylene Chloride	1.184	0.942	20.4	96	0.00
21 T	trans-1,2-Dichloroethene	0.862	0.852	1.2	96	0.00
22 T	Diisopropyl ether	3.050	3.184	-4.4	97	0.00
23 T	Vinyl Acetate	2.645	2.826	-6.8	96	0.00
24 P	1,1-Dichloroethane	1.620	1.593	1.7	96	0.00
25 T	2-Butanone	0.865	0.891	-3.0	97	0.00
26 T	2,2-Dichloropropane	1.172	1.026	12.5	89	0.00
27 T	cis-1,2-Dichloroethene	1.014	1.009	0.5	97	0.00
28 T	Bromochloromethane	0.685	0.721	-5.3	105	0.00
29 T	Tetrahydrofuran	0.582	0.596	-2.4	97	0.00
30 C	Chloroform	1.604	1.625	-1.3#	98	0.00
31 T	Cyclohexane	1.459	1.475	-1.1	97	0.00
32 T	1,1,1-Trichloroethane	1.304	1.324	-1.5	97	0.00
33 S	1,2-Dichloroethane-d4	1.005	1.058	-5.3	100	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	101	0.00
35 S	Dibromofluoromethane	0.478	0.471	1.5	98	0.00
36 T	1,1-Dichloropropene	0.688	0.645	6.2	95	0.00
37 T	Ethyl Acetate	0.974	0.916	6.0	96	0.00
38 T	Carbon Tetrachloride	0.642	0.608	5.3	95	0.00
39 T	Methylcyclohexane	0.888	0.822	7.4	97	0.00
40 TM	Benzene	2.164	2.050	5.3	95	0.00
41 T	Methacrylonitrile	0.522	0.483	7.5	92	0.00
42 TM	1,2-Dichloroethane	0.748	0.701	6.3	95	0.00
43 T	Isopropyl Acetate	1.428	1.403	1.8	98	0.00
44 TM	Trichloroethene	0.591	0.530	10.3	95	0.00
45 C	1,2-Dichloropropane	0.577	0.546	5.4#	97	0.00
46 T	Dibromomethane	0.393	0.361	8.1	96	0.00
47 T	Bromodichloromethane	0.714	0.676	5.3	95	0.00
48 T	Methyl methacrylate	0.703	0.683	2.8	96	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072522\
 Data File : VX030202.D
 Acq On : 25 Jul 2022 19:59
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 26 01:38:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 25 16:49:17 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.016	0.015	6.3	98	0.00
50 S	Toluene-d8	1.851	1.878	-1.5	99	0.00
51 T	4-Methyl-2-Pentanone	0.969	0.955	1.4	98	0.00
52 CM	Toluene	1.320	1.257	4.8#	94	0.00
53 T	t-1,3-Dichloropropene	0.769	0.749	2.6	96	0.00
54 T	cis-1,3-Dichloropropene	0.838	0.802	4.3	95	0.00
55 T	1,1,2-Trichloroethane	0.562	0.548	2.5	98	0.00
56 T	Ethyl methacrylate	0.889	0.897	-0.9	97	0.00
57 T	1,3-Dichloropropane	0.928	0.884	4.7	97	0.00
58 T	2-Chloroethyl Vinyl ether	0.441	0.475	-7.7	99	0.00
59 T	2-Hexanone	0.765	0.763	0.3	97	0.00
60 T	Dibromochloromethane	0.520	0.517	0.6	96	0.00
61 T	1,2-Dibromoethane	0.587	0.563	4.1	96	0.00
62 S	4-Bromofluorobenzene	0.687	0.723	-5.2	101	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	101	0.00
64 T	Tetrachloroethene	0.538	0.486	9.7	93	0.00
65 PM	Chlorobenzene	1.358	1.288	5.2	97	0.00
66 T	1,1,1,2-Tetrachloroethane	0.462	0.455	1.5	96	0.00
67 C	Ethyl Benzene	2.659	2.358	11.3#	97	0.00
68 T	m/p-Xylenes	0.920	0.898	2.4	96	0.00
69 T	o-Xylene	0.907	0.890	1.9	96	0.00
70 T	Styrene	1.490	1.513	-1.5	97	0.00
71 P	Bromoform	0.339	0.351	-3.5	97	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
73 T	Isopropylbenzene	3.903	3.700	5.2	97	0.00
74 T	N-amyl acetate	1.897	1.933	-1.9	99	0.00
75 P	1,1,2,2-Tetrachloroethane	1.429	1.336	6.5	98	0.00
76 T	1,2,3-Trichloropropane	1.318	1.258	4.6	100	0.00
77 T	Bromobenzene	0.916	0.840	8.3	94	0.00
78 T	n-propylbenzene	4.659	4.489	3.6	96	0.00
79 T	2-Chlorotoluene	2.942	2.679	8.9	97	0.00
80 T	1,3,5-Trimethylbenzene	3.268	3.153	3.5	97	0.00
81 T	trans-1,4-Dichloro-2-butene	0.439	0.421	4.1	92	0.00
82 T	4-Chlorotoluene	3.214	3.078	4.2	97	0.00
83 T	tert-Butylbenzene	3.029	2.931	3.2	98	0.00
84 T	1,2,4-Trimethylbenzene	3.261	3.163	3.0	95	0.00
85 T	sec-Butylbenzene	4.146	4.014	3.2	96	0.00
86 T	p-Isopropyltoluene	3.317	3.265	1.6	98	0.00
87 T	1,3-Dichlorobenzene	1.782	1.652	7.3	97	0.00
88 T	1,4-Dichlorobenzene	1.814	1.662	8.4	97	0.00
89 T	n-Butylbenzene	3.088	3.175	-2.8	101	0.00
90 T	Hexachloroethane	0.571	0.569	0.4	101	0.00
91 T	1,2-Dichlorobenzene	1.749	1.686	3.6	99	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.343	0.362	-5.5	110	0.00
93 T	1,2,4-Trichlorobenzene	1.014	0.966	4.7	97	0.00
94 T	Hexachlorobutadiene	0.407	0.374	8.1	96	0.00
95 T	Naphthalene	3.962	4.059	-2.4	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072522\
Data File : VX030202.D
Acq On : 25 Jul 2022 19:59
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jul 26 01:38:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072522W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 25 16:49:17 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.040	1.002	3.7	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6