

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072523\
 Data File : VX036648.D
 Acq On : 25 Jul 2023 12:55
 Operator : JC/MD
 Sample : PB154393ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB154393ZHE#01

Quant Time: Jul 26 06:24:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	166593	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	291395	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	270045	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	121036	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	109940	44.601	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.200%
35) Dibromofluoromethane	5.385	113	93330	46.598	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.200%
50) Toluene-d8	8.647	98	351606	49.503	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.000%
62) 4-Bromofluorobenzene	11.079	95	132692	49.517	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.040%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	52192	51.025	ug/l	92
20) Methylene Chloride	2.788	84	8584	3.773	ug/l	90
25) 2-Butanone	4.562	43	35098	21.149	ug/l	96
37) Ethyl Acetate	4.715	43	9343	2.765	ug/l	97
43) Isopropyl Acetate	6.342	43	161802	27.819	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072523\
Data File : VX036648.D
Acq On : 25 Jul 2023 12:55
Operator : JC/MD
Sample : PB154393ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB154393ZHE#01

Quant Time: Jul 26 06:24:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

