

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072523\
 Data File : VX036650.D
 Acq On : 25 Jul 2023 13:42
 Operator : JC/MD
 Sample : PB154393ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB154393ZHE#03

Quant Time: Jul 26 06:24:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	165091	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	292778	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	268181	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	119379	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	110746	45.336	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.680%
35) Dibromofluoromethane	5.391	113	94014	46.718	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.440%
50) Toluene-d8	8.647	98	349520	48.977	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.960%
62) 4-Bromofluorobenzene	11.079	95	132329	49.148	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.300%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	52908	52.195	ug/l	91
20) Methylene Chloride	2.788	84	11236	4.984	ug/l	94
25) 2-Butanone	4.568	43	33233	20.207	ug/l	95
37) Ethyl Acetate	4.721	43	8853	2.608	ug/l #	95
43) Isopropyl Acetate	6.342	43	156028	26.699	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072523\
Data File : VX036650.D
Acq On : 25 Jul 2023 13:42
Operator : JC/MD
Sample : PB154393ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB154393ZHE#03

Quant Time: Jul 26 06:24:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

