

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072523\
 Data File : VX036658.D
 Acq On : 25 Jul 2023 17:13
 Operator : JC/MD
 Sample : 03750-12 1000X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-22

Quant Time: Jul 26 06:27:43 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	159592	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	283998	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	262604	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	117903	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	107075	45.344	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.680%
35) Dibromofluoromethane	5.385	113	91508	46.878	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.760%
50) Toluene-d8	8.647	98	341300	49.304	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.600%
62) 4-Bromofluorobenzene	11.079	95	129371	49.535	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.080%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	1377	0.705	ug/l	91
27) cis-1,2-Dichloroethene	4.489	96	2771	1.168	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072523\
Data File : VX036658.D
Acq On : 25 Jul 2023 17:13
Operator : JC/MD
Sample : 03750-12 1000X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-22

Quant Time: Jul 26 06:27:43 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

