

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072523\
 Data File : VX036660.D
 Acq On : 25 Jul 2023 18:01
 Operator : JC/MD
 Sample : 03750-05 2000X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-1

Quant Time: Jul 26 06:28:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	161719	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	287146	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	257052	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	110242	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	108932	45.524	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.040%
35) Dibromofluoromethane	5.391	113	91403	46.311	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.620%
50) Toluene-d8	8.647	98	341803	48.835	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.680%
62) 4-Bromofluorobenzene	11.079	95	123294	46.691	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.380%
Target Compounds						
27) cis-1,2-Dichloroethene	4.489	96	143727	59.791	ug/l	84
44) Trichloroethene	7.129	130	59695	26.683	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072523\
Data File : VX036660.D
Acq On : 25 Jul 2023 18:01
Operator : JC/MD
Sample : 03750-05 2000X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-1

Quant Time: Jul 26 06:28:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

