

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072525\
 Data File : VX047130.D
 Acq On : 25 Jul 2025 10:29
 Operator : JC/MD
 Sample : VX0725WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0725WBL01

Quant Time: Jul 26 00:22:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	437780	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	806824	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	773720	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	398499	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	293706	53.858	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.720%	
35) Dibromofluoromethane	5.397	113	240871	48.352	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 96.700%	
50) Toluene-d8	8.647	98	863379	53.517	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 107.040%	
62) 4-Bromofluorobenzene	11.079	95	361638	52.017	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 104.040%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072525\
Data File : VX047130.D
Acq On : 25 Jul 2025 10:29
Operator : JC/MD
Sample : VX0725WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0725WBL01

Quant Time: Jul 26 00:22:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 22 03:59:51 2025
Response via : Initial Calibration

