

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072525\
 Data File : VX047138.D
 Acq On : 25 Jul 2025 13:25
 Operator : JC/MD
 Sample : Q2693-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Jul 26 00:29:32 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	450619	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	834470	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	801303	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	410899	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	309700	55.173	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.340%
35) Dibromofluoromethane	5.404	113	253615	49.224	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.440%
50) Toluene-d8	8.653	98	916941	54.954	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	109.900%
62) 4-Bromofluorobenzene	11.079	95	386739	53.784	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.560%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072525\
Data File : VX047138.D
Acq On : 25 Jul 2025 13:25
Operator : JC/MD
Sample : Q2693-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Jul 26 00:29:32 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 22 03:59:51 2025
Response via : Initial Calibration

