

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072525\
 Data File : VX047138.D
 Acq On : 25 Jul 2025 13:25
 Operator : JC/MD
 Sample : Q2693-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Jul 26 00:29:32 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	450619	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	834470	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	801303	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	410899	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	309700	55.173	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.340%
35) Dibromofluoromethane	5.404	113	253615	49.224	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.440%
50) Toluene-d8	8.653	98	916941	54.954	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	109.900%
62) 4-Bromofluorobenzene	11.079	95	386739	53.784	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.560%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX072525\
Data File : VX047138.D
Acq On : 25 Jul 2025 13:25
Operator : JC/MD
Sample : Q2693-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Jul 26 00:29:32 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 22 03:59:51 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.568 min Scan# 71
Delta R.T. 0.006 min
Lab File: VX047138.D
Acq: 25 Jul 2025 13:25




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4


Tgt Ion:168 Resp: 450619
Ion Ratio Lower Upper
168 100
99 61.5 48.0 72.0



#33
1,2-Dichloroethane-d4
Concen: 55.173 ug/l
RT: 5.964 min Scan# 800
Delta R.T. 0.000 min
Lab File: VX047138.D
Acq: 25 Jul 2025 13:25

Tgt Ion: 65 Resp: 309700
Ion Ratio Lower Upper
65 100
67 51.4 0.0 104.8









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.769 min Scan# 911
Delta R.T. 0.000 min
Lab File: VX047138.D
Acq: 25 Jul 2025 13:25




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4


Tgt Ion:114 Resp: 834470
Ion Ratio Lower Upper
114 100
63 21.8 0.0 43.2
88 16.5 0.0 33.8



#35
Dibromofluoromethane
Concen: 49.224 ug/l
RT: 5.404 min Scan# 708
Delta R.T. 0.007 min
Lab File: VX047138.D
Acq: 25 Jul 2025 13:25

Tgt Ion:113 Resp: 253615
Ion Ratio Lower Upper
113 100
111 102.5 82.6 124.0
192 18.0 14.9 22.3









#50
Toluene-d8
Concen: 54.954 ug/l
RT: 8.653 min Scan# 11
Delta R.T. 0.001 min
Lab File: VX047138.D
Acq: 25 Jul 2025 13:25




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4


Tgt Ion: 98 Resp: 916941
Ion Ratio Lower Upper
98 100
100 66.3 53.3 79.9



#62
4-Bromofluorobenzene
Concen: 53.784 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX047138.D
Acq: 25 Jul 2025 13:25

Tgt Ion: 95 Resp: 386739
Ion Ratio Lower Upper
95 100
174 72.5 0.0 144.6
176 67.8 0.0 139.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX047138.D
Acq: 25 Jul 2025 13:25

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Tgt Ion:117 Resp: 801303
Ion Ratio Lower Upper
117 100
82 56.3 45.4 68.2
119 32.6 26.1 39.1



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.018 min Scan# 1793
Delta R.T. 0.001 min
Lab File: VX047138.D
Acq: 25 Jul 2025 13:25

Tgt Ion:152 Resp: 410899
Ion Ratio Lower Upper
152 100
115 62.9 45.6 136.7
150 157.9 0.0 354.6

