

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072618\
 Data File : VX003673.D
 Acq On : 26 Jul 2018 16:37
 Operator : JC/SP
 Sample : J4104-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LC-SD-EB03-180719

Quant Time: Jul 27 13:32:45 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072418W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 27 02:16:51 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	91834	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	133182	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	123881	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	63910	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	76147	55.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	111.96%	
35) Dibromofluoromethane	5.50	113	62240	57.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	114.34%	
50) Toluene-d8	8.72	98	217242	52.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.56%	
62) 4-Bromofluorobenzene	11.14	95	72558	47.88	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.76%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072618\
Data File : VX003673.D
Acq On : 26 Jul 2018 16:37
Operator : JC/SP
Sample : J4104-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LC-SD-EB03-180719

Quant Time: Jul 27 13:32:45 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072418W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 27 02:16:51 2018
Response via : Initial Calibration

