

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072618\
 Data File : VX003691.D
 Acq On : 27 Jul 2018 02:00
 Operator : JC/SP
 Sample : J4154-11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FL-EB-2018-1

Quant Time: Jul 27 06:56:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072418W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 27 02:16:51 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	82398	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	115713	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	121244	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	59224	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	68089	55.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.58%	
35) Dibromofluoromethane	5.51	113	59021	62.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	124.78%	
50) Toluene-d8	8.72	98	213385	59.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.22%	
62) 4-Bromofluorobenzene	11.14	95	68006	51.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.30%	
Target Compounds						
16) Acetone	2.45	43	7358	12.83	ug/l	97
20) Methylene Chloride	2.86	84	35202	30.98	ug/l	96
25) 2-Butanone	4.71	43	13458	15.19	ug/l	99
29) Tetrahydrofuran	5.17	42	216360	375.21	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072618\
Data File : VX003691.D
Acq On : 27 Jul 2018 02:00
Operator : JC/SP
Sample : J4154-11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FL-EB-2018-1

Quant Time: Jul 27 06:56:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072418W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 27 02:16:51 2018
Response via : Initial Calibration

