

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072618\
 Data File : VX003692.D
 Acq On : 27 Jul 2018 02:26
 Operator : JC/SP
 Sample : J4154-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FL-TB-2018-1

Quant Time: Jul 27 17:21:05 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072418W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 27 02:16:51 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	92914	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	141942	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	133622	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	68001	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	77011	55.95	ug/l	0.00
Spiked Amount	50.000		Recovery	=	111.90%	
35) Dibromofluoromethane	5.51	113	62768	54.09	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.18%	
50) Toluene-d8	8.72	98	229127	51.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.48%	
62) 4-Bromofluorobenzene	11.14	95	75737	46.89	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.78%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072618\
Data File : VX003692.D
Acq On : 27 Jul 2018 02:26
Operator : JC/SP
Sample : J4154-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FL-TB-2018-1

Quant Time: Jul 27 17:21:05 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072418W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 27 02:16:51 2018
Response via : Initial Calibration

