

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072618\
 Data File : VX003697.D
 Acq On : 27 Jul 2018 04:40
 Operator : JC/SP
 Sample : J4154-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FL-0508

Quant Time: Jul 27 17:35:35 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072418W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 27 02:16:51 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	87298	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	154559	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	136576	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	74939	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	74666	57.74	ug/l	0.00
Spiked Amount			Recovery	=	115.48%	
35) Dibromofluoromethane	5.50	113	57649	45.63	ug/l	0.00
Spiked Amount			Recovery	=	91.26%	
50) Toluene-d8	8.72	98	247087	51.24	ug/l	0.00
Spiked Amount			Recovery	=	102.48%	
62) 4-Bromofluorobenzene	11.14	95	90969	51.72	ug/l	0.00
Spiked Amount			Recovery	=	103.44%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
31) Cyclohexane	5.58	56	31713	18.24	ug/l	99
39) Methylcyclohexane	7.46	83	58771	29.55	ug/l	98
40) Benzene	6.14	78	39809	7.45	ug/l	98
67) Ethyl Benzene	10.26	91	227569	39.55	ug/l	100
68) m/p-Xylenes	10.36	106	189366	85.79	ug/l	99
69) o-Xylene	10.70	106	1437	0.68	ug/l	97
73) Isopropylbenzene	11.02	105	23378	4.57	ug/l	99
78) n-propylbenzene	11.36	91	27594	4.67	ug/l	98
80) 1,3,5-Trimethylbenzene	11.51	105	36481	8.42	ug/l	99
84) 1,2,4-Trimethylbenzene	11.81	105	92854	20.86	ug/l	99
85) sec-Butylbenzene	11.95	105	4992	0.99	ug/l	98
86) p-Isopropyltoluene	12.07	119	4542	1.00	ug/l	93
89) n-Butylbenzene	12.38	91	4278	1.05	ug/l #	35
95) Naphthalene	13.83	128	24961	4.37	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072618\
Data File : VX003697.D
Acq On : 27 Jul 2018 04:40
Operator : JC/SP
Sample : J4154-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FL-0508

Quant Time: Jul 27 17:35:35 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072418W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 27 02:16:51 2018
Response via : Initial Calibration

