

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072618\
 Data File : VX003703.D
 Acq On : 27 Jul 2018 07:20
 Operator : JC/SP
 Sample : J4100-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY030LC026-W-180719

Quant Time: Jul 27 17:37:36 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072418W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 27 02:16:51 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	80970	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	136754	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	135227	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	70831	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	64150	53.49	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.98%	
35) Dibromofluoromethane	5.51	113	52991	47.40	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.80%	
50) Toluene-d8	8.72	98	218910	51.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.62%	
62) 4-Bromofluorobenzene	11.14	95	72126	46.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.70%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072618\
Data File : VX003703.D
Acq On : 27 Jul 2018 07:20
Operator : JC/SP
Sample : J4100-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY030LC026-W-180719

Quant Time: Jul 27 17:37:36 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072418W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 27 02:16:51 2018
Response via : Initial Calibration

