

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072618\
 Data File : VX003705.D
 Acq On : 27 Jul 2018 08:13
 Operator : JC/SP
 Sample : J4073-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE117D1-071618

Quant Time: Jul 27 17:38:34 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072418W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 27 02:16:51 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	87357	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	134216	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	134861	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	79444	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	71912	55.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.14%	
35) Dibromofluoromethane	5.51	113	59309	54.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.12%	
50) Toluene-d8	8.72	98	235768	56.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.62%	
62) 4-Bromofluorobenzene	11.14	95	70940	46.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.90%	
Target Compounds						
44) Trichloroethene	7.21	130	13197	10.60	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072618\
Data File : VX003705.D
Acq On : 27 Jul 2018 08:13
Operator : JC/SP
Sample : J4073-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE117D1-071618

Quant Time: Jul 27 17:38:34 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072418W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 27 02:16:51 2018
Response via : Initial Calibration

