

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072618\
 Data File : VX003707.D
 Acq On : 27 Jul 2018 09:06
 Operator : JC/SP
 Sample : J4073-16
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE108D1-071718RE

Quant Time: Jul 27 10:55:27 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072418W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 25 04:33:52 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	84280	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	133625	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	132680	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	72288	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	74376	59.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.16%	
35) Dibromofluoromethane	5.51	113	56738	51.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.88%	
50) Toluene-d8	8.72	98	236087	56.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.26%	
62) 4-Bromofluorobenzene	11.14	95	86638	56.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.96%	
Target Compounds						
44) Trichloroethene	7.21	130	27156	21.91	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072618\
Data File : VX003707.D
Acq On : 27 Jul 2018 09:06
Operator : JC/SP
Sample : J4073-16
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE108D1-071718RE

Quant Time: Jul 27 10:55:27 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072418W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 25 04:33:52 2018
Response via : Initial Calibration

