

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072618\
 Data File : VX003708.D
 Acq On : 27 Jul 2018 09:33
 Operator : JC/SP
 Sample : J4114-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY030LC058-W-180720

Quant Time: Jul 27 10:57:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072418W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 25 04:33:52 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	84505	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	139551	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	132631	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	82514	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	69078	55.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.36%	
35) Dibromofluoromethane	5.51	113	55183	48.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.74%	
50) Toluene-d8	8.72	98	219246	50.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.72%	
62) 4-Bromofluorobenzene	11.14	95	85275	53.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.40%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072618\
Data File : VX003708.D
Acq On : 27 Jul 2018 09:33
Operator : JC/SP
Sample : J4114-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY030LC058-W-180720

Quant Time: Jul 27 10:57:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072418W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 25 04:33:52 2018
Response via : Initial Calibration

