

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072618\
 Data File : VX003681.D
 Acq On : 26 Jul 2018 20:12
 Operator : JC/SP
 Sample : J4069-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY0300F003-180718

Quant Time: Jul 27 13:38:57 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072418W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 27 02:16:51 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	82048	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	114664	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	112680	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	63843	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	61482	50.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.18%	
35) Dibromofluoromethane	5.51	113	56672	60.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	120.92%	
50) Toluene-d8	8.72	98	202264	56.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.08%	
62) 4-Bromofluorobenzene	11.14	95	63155	48.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.80%	


Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072618\
Data File : VX003681.D
Acq On : 26 Jul 2018 20:12
Operator : JC/SP
Sample : J4069-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY0300F003-180718

Quant Time: Jul 27 13:38:57 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072418W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 27 02:16:51 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 5.67 min Scan# 833

Delta R.T. 0.00 min

Lab File: VX003681.D

Acq: 26 Jul 2018 20:12

Instrument :

MSVOA_X

ClientSampleId :


KY0300F003-180718

Tot Ion:168 Resp: 82048

Ion Ratio Lower Upper




168 100

99 54.6 49.5 74.3



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#33

1,2-Dichloroethane-d4

Concen: 50.59 ug/l

RT: 6.07 min Scan# 900

Delta R.T. 0.00 min

Lab File: VX003681.D


Acq: 26 Jul 2018 20:12

Tgt Ion: 65 Resp: 61482

Ion Ratio Lower Upper



65 100


67 53.3 0.0 107.8



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0





#34

1,4-Difluorobenzene

Concen: 50.00 ug/l

RT: 6.87 min Scan# 1030

Delta R.T. 0.00 min

Lab File: VX003681.D

Acq: 26 Jul 2018 20:12

Instrument :

MSVOA_X

ClientSampleId :

KY030OF003-180718


Tot Ion:114 Resp: 114664

Ion Ratio Lower Upper

114 100

63 21.0 0.0 41.0




88 15.6 0.0 31.8



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 60.46 ug/l

RT: 5.51 min Scan# 807

Delta R.T. 0.00 min

Lab File: VX003681.D

Acq: 26 Jul 2018 20:12


Tgt Ion:113 Resp: 56672

Ion Ratio Lower Upper

113 100

111 103.1 81.2 121.8



192 22.0 18.2 27.4




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#50

Toluene-d8

Concen: 56.54 ug/l

RT: 8.72 min Scan# 1334

Delta R.T. 0.00 min

Lab File: VX003681.D

Acq: 26 Jul 2018 20:12

Instrument :

MSVOA_X

ClientSampleId :


KY0300F003-180718

Tot Ion: 98 Resp: 202264

Ion Ratio Lower Upper

98 100

100 63.0 50.9 76.3



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): V

150000

100000



50000

0

8.60 8.70 8.80 8.90

8.72

Time-->



#62

4-Bromofluorobenzene

Concen: 48.40 ug/l

RT: 11.14 min Scan# 1731

Delta R.T. 0.00 min

Lab File: VX003681.D

Acq: 26 Jul 2018 20:12


Tgt Ion: 95 Resp: 63155

Ion Ratio Lower Upper

95 100

174 92.5 0.0 179.0

176 89.3 0.0 173.0



Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V

60000

40000


20000


0

11.05 11.10 11.15 11.20

11.14

Time-->









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.12 min Scan# 1563
Delta R.T. 0.00 min
Lab File: VX003681.D
Acq: 26 Jul 2018 20:12

Instrument :
MSVOA_X
ClientSampleId :
KY0300F003-180718

Tot Ion:117 Resp: 112680
Ion Ratio Lower Upper
117 100
82 54.5 44.3 66.5
119 32.0 25.6 38.4




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.08 min Scan# 1885
Delta R.T. -0.01 min
Lab File: VX003681.D
Acq: 26 Jul 2018 20:12

Tgt Ion:152 Resp: 63843
Ion Ratio Lower Upper
152 100
115 56.1 40.0 119.9
150 156.3 0.0 348.4



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

