

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072618\
 Data File : VX003708.D
 Acq On : 27 Jul 2018 09:33
 Operator : JC/SP
 Sample : J4114-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY030LC058-W-180720

Quant Time: Jul 27 10:57:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X072418W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 25 04:33:52 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	84505	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	139551	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	132631	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	82514	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	69078	55.18	ug/l	0.00
Spiked Amount			Recovery	=	110.36%	
35) Dibromofluoromethane	5.51	113	55183	48.37	ug/l	0.00
Spiked Amount			Recovery	=	96.74%	
50) Toluene-d8	8.72	98	219246	50.36	ug/l	0.00
Spiked Amount			Recovery	=	100.72%	
62) 4-Bromofluorobenzene	11.14	95	85275	53.70	ug/l	0.00
Spiked Amount			Recovery	=	107.40%	


Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072618\
Data File : VX003708.D
Acq On : 27 Jul 2018 09:33
Operator : JC/SP
Sample : J4114-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY030LC058-W-180720

Quant Time: Jul 27 10:57:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X072418W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 25 04:33:52 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 5.67 min Scan# 833

Delta R.T. 0.00 min

Lab File: VX003708.D

Acq: 27 Jul 2018 09:33

Instrument :

MSVOA_X

ClientSampleId :


KY030LC058-W-180720

Tot Ion:168 Resp: 84505

Ion Ratio Lower Upper




168 100

99 57.8 49.5 74.3



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#33

1,2-Dichloroethane-d4

Concen: 55.18 ug/l

RT: 6.07 min Scan# 899

Delta R.T. -0.01 min

Lab File: VX003708.D


Acq: 27 Jul 2018 09:33

Tgt Ion: 65 Resp: 69078

Ion Ratio Lower Upper



65 100


67 52.8 0.0 107.8



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0









#34
 1,4-Difluorobenzene
 Concen: 50.00 ug/l
 RT: 6.87 min Scan# 1030
 Delta R.T. 0.00 min
 Lab File: VX003708.D
 Acq: 27 Jul 2018 09:33

Instrument :
 MSVOA_X
 ClientSampleId :
 KY030LC058-W-180720

Tot Ion:114 Resp: 139551
 Ion Ratio Lower Upper
 114 100
 63 22.8 0.0 41.0
 88 15.1 0.0 31.8




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VX0
 Ion 87.90 (87.60 to 88.60): VX0






#35
 Dibromofluoromethane
 Concen: 48.37 ug/l
 RT: 5.51 min Scan# 807
 Delta R.T. 0.00 min
 Lab File: VX003708.D
 Acq: 27 Jul 2018 09:33

Tgt Ion:113 Resp: 55183
 Ion Ratio Lower Upper
 113 100
 111 102.2 81.2 121.8
 192 21.3 18.2 27.4



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 50.36 ug/l
RT: 8.72 min Scan# 1334
Delta R.T. 0.00 min
Lab File: VX003708.D
Acq: 27 Jul 2018 09:33




Instrument :
MSVOA_X
ClientSampleId :
KY030LC058-W-180720


Tot Ion: 98 Resp: 219246
Ion Ratio Lower Upper
98 100
100 63.9 50.9 76.3



#62
4-Bromofluorobenzene
Concen: 53.70 ug/l
RT: 11.14 min Scan# 1731
Delta R.T. 0.00 min
Lab File: VX003708.D
Acq: 27 Jul 2018 09:33

Tgt Ion: 95 Resp: 85275
Ion Ratio Lower Upper
95 100
174 93.8 0.0 179.0
176 90.9 0.0 173.0









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.12 min Scan# 1563
Delta R.T. 0.00 min
Lab File: VX003708.D
Acq: 27 Jul 2018 09:33

Instrument :
MSVOA_X
ClientSampleId :
KY030LC058-W-180720

Tot Ion:117 Resp: 132631
Ion Ratio Lower Upper
117 100
82 54.9 44.3 66.5
119 32.1 25.6 38.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.08 min Scan# 1885
Delta R.T. -0.01 min
Lab File: VX003708.D
Acq: 27 Jul 2018 09:33

Tgt Ion:152 Resp: 82514
Ion Ratio Lower Upper
152 100
115 55.6 40.0 119.9
150 156.7 0.0 348.4

