

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072619\
 Data File : VX011125.D
 Acq On : 27 Jul 2019 05:50
 Operator : JC/SP
 Sample : K3620-18
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CL-02-072519-D

Quant Time: Jul 29 01:39:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 24 05:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	194780	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	317771	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	305110	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	141089	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	107229	51.75	ug/l	0.00
Spiked Amount			Recovery	=	103.50%	
35) Dibromofluoromethane	5.50	113	97974	48.58	ug/l	0.00
Spiked Amount			Recovery	=	97.16%	
50) Toluene-d8	8.71	98	389521	51.49	ug/l	0.00
Spiked Amount			Recovery	=	102.98%	
62) 4-Bromofluorobenzene	11.13	95	137631	49.50	ug/l	0.00
Spiked Amount			Recovery	=	99.00%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	1.21	85	23	0.557	ug/l #	43
5) Bromomethane	1.65	94	356	0.346	ug/l #	61
6) Chloroethane	1.62	64	578	0.498	ug/l #	41
8) Diethyl Ether	2.19	74	316	0.297	ug/l #	2
11) Tert butyl alcohol	3.05	59	856	1.774	ug/l #	73
16) Acetone	2.45	43	10813	11.554	ug/l	95
18) Methyl Acetate	2.77	43	1597	0.749	ug/l	98
20) Methylene Chloride	2.86	84	11124	5.609	ug/l	90
25) 2-Butanone	4.68	43	4581	3.410	ug/l	100
29) Tetrahydrofuran	5.15	42	574	0.675	ug/l #	81
30) Chloroform	5.21	83	741	0.222	ug/l	95
36) 1,1-Dichloropropene	5.66	75	13395	5.132	ug/l #	49
38) Carbon Tetrachloride	5.66	117	18195	6.689	ug/l #	18
43) Isopropyl Acetate	6.45	43	10585	2.409	ug/l	97
48) Methyl methacrylate	7.68	41	5273	2.463	ug/l #	10
49) 1,4-Dioxane	7.75	88	24	0.401	ug/l #	28
51) 4-Methyl-2-Pentanone	8.61	43	52253	18.341	ug/l #	48
54) cis-1,3-Dichloropropene	8.61	75	26893	8.683	ug/l #	60
57) 1,3-Dichloropropane	9.54	76	6313	1.840	ug/l #	45
59) 2-Hexanone	9.54	43	75169	34.018	ug/l #	51
60) Dibromochloromethane	9.59	129	74	2.556	ug/l #	11
74) N-amyl acetate	10.83	43	989	0.272	ug/l #	50
76) 1,2,3-Trichloropropane	11.13	75	68165	26.773	ug/l #	35
81) trans-1,4-Dichloro-2-buten	11.13	75	68165	66.412	ug/l #	11
95) Naphthalene	13.83	128	2881	0.318	ug/l #	93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072619\
Data File : VX011125.D
Acq On : 27 Jul 2019 05:50
Operator : JC/SP
Sample : K3620-18
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleID :
CL-02-072519-D

Quant Time: Jul 29 01:39:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 24 05:39:23 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.67 min Scan# 740

Delta R.T. 0.01 min

Lab File: VX011125.D

Acq: 27 Jul 2019 05:50

Instrument :

MSVOA_X

ClientSampleId :


CL-02-072519-D

Tot Ion:168 Resp: 194780

Ion Ratio Lower Upper

168 100

99 50.6 37.6 56.4



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.67

80000



60000

40000

20000

0

Time--> 5.50 5.60 5.70 5.80



#2

Dichlorodifluoromethane

Concen: 0.557 ug/l

RT: 1.21 min Scan# 9

Delta R.T. 0.01 min

Lab File: VX011125.D


Acq: 27 Jul 2019 05:50

Tgt Ion: 85 Resp: 23

Ion Ratio Lower Upper

85 100

87 0.0 15.9 47.7#



Abundance Ion 84.90 (84.60 to 85.60): VX0

Ion 86.90 (86.60 to 87.60): VX0

1.21


60


40

20

0

Time--> 1.19 1.20 1.21 1.22





#5

Bromomethane

Concen: 0.346 ug/l

RT: 1.65 min Scan# 81

Delta R.T. 0.01 min

Lab File: VX011125.D

Acq: 27 Jul 2019 05:50

Instrument :

MSVOA_X

ClientSampleId :


CL-02-072519-D

Tot Ion: 94 Resp: 356

Ion Ratio Lower Upper




94 100

96 56.8 75.6 113.4#



Abundance Ion 93.90 (93.60 to 94.60): VX0

Ion 95.90 (95.60 to 96.60): VX0



#6

Chloroethane

Concen: 0.498 ug/l

RT: 1.62 min Scan# 77

Delta R.T. -0.10 min

Lab File: VX011125.D


Acq: 27 Jul 2019 05:50

Tgt Ion: 64 Resp: 578

Ion Ratio Lower Upper



64 100


66 0.0 27.0 40.6#



Abundance Ion 63.90 (63.60 to 64.60): VX0

Ion 65.90 (65.60 to 66.60): VX0





#8

Diethyl Ether

Concen: 0.297 ug/l

RT: 2.19 min Scan# 170

Delta R.T. -0.00 min

Lab File: VX011125.D

Acq: 27 Jul 2019 05:50

Instrument :

MSVOA_X

ClientSampleId :





CL-02-072519-D

Tot Ion: 74 Resp: 316

Ion Ratio Lower Upper

74 100

45 184.2 45.5 136.5#



#11

Tert butyl alcohol

Concen: 1.774 ug/l

RT: 3.05 min Scan# 311

Delta R.T. -0.00 min

Lab File: VX011125.D




Acq: 27 Jul 2019 05:50


Tgt Ion: 59 Resp: 856

Ion Ratio Lower Upper

59 100

57 0.0 7.9 11.9#





#16

Acetone

Concen: 11.554 ug/l

RT: 2.45 min Scan# 212

Delta R.T. -0.00 min

Lab File: VX011125.D

Acq: 27 Jul 2019 05:50

Instrument :

MSVOA_X

ClientSampleId :


CL-02-072519-D

Tot Ion: 43 Resp: 10813

Ion Ratio Lower Upper




43 100

58 31.5 27.7 41.5



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0



#18

Methyl Acetate

Concen: 0.749 ug/l

RT: 2.77 min Scan# 265

Delta R.T. -0.01 min

Lab File: VX011125.D


Acq: 27 Jul 2019 05:50

Tgt Ion: 43 Resp: 1597

Ion Ratio Lower Upper



43 100


74 25.7 21.5 32.3



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 74.00 (73.70 to 74.70): VX0






#20
Methvlene Chloride
Concen: 5.609 ug/l
RT: 2.86 min Scan# 280
Delta R.T. 0.01 min
Lab File: VX011125.D
Acq: 27 Jul 2019 05:50

Instrument :
MSVOA_X
ClientSampleId :
CL-02-072519-D

Tot Ion:	84	Resp:	11124
Ion	Ratio	Lower	Upper
84	100		
49	97.3	89.9	134.9
51	32.3	27.8	41.6
86	59.7	51.8	77.6






Abundance

Ion 83.90 (83.60 to 84.60): VX0

Ion 48.90 (48.60 to 49.60): VX0


Ion 50.90 (50.60 to 51.60): VX0

Ion 85.90 (85.60 to 86.60): VX0



#25
2-Butanone
Concen: 3.410 ug/l
RT: 4.68 min Scan# 578
Delta R.T. 0.01 min
Lab File: VX011125.D
Acq: 27 Jul 2019 05:50



Tgt Ion:	43	Resp:	4581
Ion	Ratio	Lower	Upper
43	100		
72	28.6	23.0	34.6




Abundance

Ion 43.00 (42.70 to 43.70): VX0

Ion 72.00 (71.70 to 72.70): VX0





#29

Tetrahydrofuran

Concen: 0.675 ug/l

RT: 5.15 min Scan# 656

Delta R.T. 0.02 min

Lab File: VX011125.D

Acq: 27 Jul 2019 05:50

Instrument :

MSVOA_X

ClientSampled :

CL-02-072519-D





Tot Ion: 42 Resp: 574

Ion Ratio Lower Upper

42 100

72 34.5 39.8 59.8#

71 56.8 37.0 55.4#



#30

Chloroform

Concen: 0.222 ug/l

RT: 5.21 min Scan# 665

Delta R.T. -0.00 min

Lab File: VX011125.D




Acq: 27 Jul 2019 05:50


Tgt Ion: 83 Resp: 741

Ion Ratio Lower Upper

83 100

85 68.3 51.4 77.0









#33
 1,2-Dichloroethane-d4
 Concen: 51.747 ug/l
 RT: 6.06 min Scan# 805
 Delta R.T. -0.00 min
 Lab File: VX011125.D
 Acq: 27 Jul 2019 05:50




Instrument :
 MSVOA_X
 ClientSampleId :
 CL-02-072519-D


Tot Ion: 65 Resp: 107229
 Ion Ratio Lower Upper
 65 100
 67 54.0 0.0 107.8



#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.86 min Scan# 936
 Delta R.T. -0.00 min
 Lab File: VX011125.D
 Acq: 27 Jul 2019 05:50

Tgt Ion: 114 Resp: 317771
 Ion Ratio Lower Upper
 114 100
 63 18.0 0.0 35.6
 88 14.8 0.0 29.8









#35
 Dibromofluoromethane
 Concen: 48.575 ug/l
 RT: 5.50 min Scan# 713
 Delta R.T. -0.00 min
 Lab File: VX011125.D
 Acq: 27 Jul 2019 05:50




Instrument :
 MSVOA_X
 ClientSampleId :
 CL-02-072519-D


Tot Ion:113 Resp: 97974
 Ion Ratio Lower Upper
 113 100
 111 102.1 82.3 123.5
 192 21.7 17.4 26.2



#36
 1,1-Dichloropropene
 Concen: 5.132 ug/l
 RT: 5.66 min Scan# 739
 Delta R.T. -0.14 min
 Lab File: VX011125.D
 Acq: 27 Jul 2019 05:50

Tgt Ion: 75 Resp: 13395
 Ion Ratio Lower Upper
 75 100
 110 10.7 19.9 59.7#
 77 0.0 25.0 37.4#





#38

Carbon Tetrachloride

Concen: 6.689 ug/l

RT: 5.66 min Scan# 739

Delta R.T. -0.13 min

Lab File: VX011125.D

Acq: 27 Jul 2019 05:50

Instrument :

MSVOA_X

ClientSampleId :

CL-02-072519-D


Tot Ion:117 Resp: 18195

Ion Ratio Lower Upper

117 100

119 6.0 74.2 111.2#




121 0.0 24.2 36.2#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V



#43

Isopropyl Acetate

Concen: 2.409 ug/l

RT: 6.45 min Scan# 868

Delta R.T. -0.00 min

Lab File: VX011125.D

Acq: 27 Jul 2019 05:50


Tgt Ion: 43 Resp: 10585

Ion Ratio Lower Upper

43 100

61 19.7 17.1 25.7



87 15.5 12.0 18.0




Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 61.00 (60.70 to 61.70): VX0

Ion 87.00 (86.70 to 87.70): VX0









#48
Methvl methacrylate
Concen: 2.463 ug/l
RT: 7.68 min Scan# 1070
Delta R.T. -0.09 min
Lab File: VX011125.D
Acq: 27 Jul 2019 05:50




Instrument :
MSVOA_X
ClientSampleId :
CL-02-072519-D


Tot Ion: 41 Resp: 5273
Ion Ratio Lower Upper
41 100
69 0.4 74.4 111.6#
39 0.0 48.9 73.3#



#49
1,4-Dioxane
Concen: 0.401 ug/l
RT: 7.75 min Scan# 1082
Delta R.T. 0.01 min
Lab File: VX011125.D
Acq: 27 Jul 2019 05:50

Tgt Ion: 88 Resp: 24
Ion Ratio Lower Upper
88 100
43 0.0 22.9 34.3#
58 0.0 49.8 74.8#









#50
Toluene-d8
Concen: 51.491 ug/l
RT: 8.71 min Scan# 1240
Delta R.T. -0.00 min
Lab File: VX011125.D
Acq: 27 Jul 2019 05:50




Instrument :
MSVOA_X
ClientSampleId :
CL-02-072519-D


Tot Ion: 98 Resp: 389521
Ion Ratio Lower Upper
98 100
100 63.7 50.6 75.8



#51
4-Methyl-2-Pentanone
Concen: 18.341 ug/l
RT: 8.61 min Scan# 1223
Delta R.T. -0.03 min
Lab File: VX011125.D
Acq: 27 Jul 2019 05:50

Tgt Ion: 43 Resp: 52253
Ion Ratio Lower Upper
43 100
58 8.3 32.2 48.4#






#54




cis-1,3-Dichloropropene
 Concen: 8.683 ug/l
 RT: 8.61 min Scan# 1223
 Delta R.T. 0.18 min
 Lab File: VX011125.D
 Acq: 27 Jul 2019 05:50

Instrument :
 MSVOA_X
 ClientSampleId :
 CL-02-072519-D

Tot Ion	Ratio	Lower	Upper
75	100		
77	0.8	26.1	39.1#
39	27.6	36.6	54.8#




Abundance Ion 74.90 (74.60 to 75.60): VX0
 Ion 76.90 (76.60 to 77.60): VX0
 Ion 39.00 (38.70 to 39.70): VX0





#57


1,3-Dichloropropane
 Concen: 1.840 ug/l
 RT: 9.54 min Scan# 1375
 Delta R.T. 0.17 min
 Lab File: VX011125.D
 Acq: 27 Jul 2019 05:50

Tgt Ion	Ratio	Lower	Upper
76	100		
78	1.1	25.6	38.4#



Abundance Ion 75.90 (75.60 to 76.60): VX0
 Ion 77.90 (77.60 to 78.60): VX0









#59
2-Hexanone
Concen: 34.018 ug/l
RT: 9.54 min Scan# 1375
Delta R.T. 0.05 min
Lab File: VX011125.D
Acq: 27 Jul 2019 05:50




Instrument :
MSVOA_X
ClientSampleId :
CL-02-072519-D


Tot Ion: 43 Resp: 75169
Ion Ratio Lower Upper
43 100
58 20.5 28.0 83.9#



#60
Dibromochloromethane
Concen: 2.556 ug/l
RT: 9.59 min Scan# 1383
Delta R.T. 0.01 min
Lab File: VX011125.D
Acq: 27 Jul 2019 05:50

Tgt Ion:129 Resp: 74
Ion Ratio Lower Upper
129 100
127 0.0 38.6 115.8#





#62

4-Bromofluorobenzene

Concen: 49.496 ug/l

RT: 11.13 min Scan# 1637

Delta R.T. -0.00 min

Lab File: VX011125.D

Acq: 27 Jul 2019 05:50

Instrument :

MSVOA_X

ClientSampleId :

CL-02-072519-D


Tot Ion: 95 Resp: 137631

Ion Ratio Lower Upper

95 100

174 92.0 0.0 180.6




176 88.0 0.0 173.8



Abundance Ion 94.90 (94.60 to 95.60): VX011125.D (-1588) (-)

Ion 173.80 (173.50 to 174.50): VX011125.D (-1588) (-)

Ion 175.80 (175.50 to 176.50): VX011125.D (-1588) (-)



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.11 min Scan# 1469

Delta R.T. -0.00 min

Lab File: VX011125.D

Acq: 27 Jul 2019 05:50


Tgt Ion: 117 Resp: 305110

Ion Ratio Lower Upper

117 100

82 51.8 42.6 63.8



119 32.2 25.9 38.9




Abundance Ion 116.90 (116.60 to 117.60): VX011125.D (-1420) (-)

Ion 82.00 (81.70 to 82.70): VX011125.D (-1420) (-)

Ion 118.90 (118.60 to 119.60): VX011125.D (-1420) (-)





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.07 min Scan# 1791

Delta R.T. -0.01 min

Lab File: VX011125.D

Acq: 27 Jul 2019 05:50

Instrument :

MSVOA_X

ClientSampleId :

CL-02-072519-D


Tot Ion:152 Resp: 141089

Ion Ratio Lower Upper

152 100

115 55.5 39.7 119.1




150 158.9 0.0 345.6



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#74

N-ethyl acetate

Concen: 0.272 ug/l

RT: 10.83 min Scan# 1587

Delta R.T. -0.07 min

Lab File: VX011125.D

Acq: 27 Jul 2019 05:50

Tgt Ion: 43 Resp: 989


Ion Ratio Lower Upper

43 100

70 0.0 34.7 52.1#

55 15.2 19.1 28.7#

61 0.0 19.9 29.9#






Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 70.00 (69.70 to 70.70): VX0

Ion 55.00 (54.70 to 55.70): VX0

Ion 61.00 (60.70 to 61.70): VX0









#76
 1,2,3-Trichloropropane
 Concen: 26.773 ug/l
 RT: 11.13 min Scan# 1637
 Delta R.T. -0.16 min
 Lab File: VX011125.D
 Acq: 27 Jul 2019 05:50




Instrument :
 MSVOA_X
 ClientSampleId :
 CL-02-072519-D


Tot Ion: 75 Resp: 68165
 Ion Ratio Lower Upper
 75 100
 77 1.2 21.3 64.0#



#81
 trans-1,4-Dichloro-2-butene
 Concen: 66.412 ug/l
 RT: 11.13 min Scan# 1637
 Delta R.T. 0.06 min
 Lab File: VX011125.D
 Acq: 27 Jul 2019 05:50

Tgt Ion: 75 Resp: 68165
 Ion Ratio Lower Upper
 75 100
 53 0.1 76.7 115.1#
 89 0.2 38.5 57.7#





#95
Naphthalene
Concen: 0.318 ug/l
RT: 13.83 min Scan# 2080
Delta R.T. -0.00 min
Lab File: VX011125.D
Acq: 27 Jul 2019 05:50

Instrument :
MSVOA_X
ClientSampleId :
CL-02-072519-D

Tot Ion	Ratio	Lower	Upper
128	100		
127	16.7	10.4	15.6#
129	12.7	8.9	13.3

