

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072619\
 Data File : VX011085.D
 Acq On : 26 Jul 2019 13:05
 Operator : JC/SP
 Sample : K4014-02 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2

Quant Time: Jul 27 02:34:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 24 05:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	208565	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	332395	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	305038	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	145479	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	110335	49.73	ug/l	0.00
Spiked Amount			Recovery	=	99.46%	
35) Dibromofluoromethane	5.50	113	101846	48.27	ug/l	0.00
Spiked Amount			Recovery	=	96.54%	
50) Toluene-d8	8.71	98	395243	49.95	ug/l	0.00
Spiked Amount			Recovery	=	99.90%	
62) 4-Bromofluorobenzene	11.13	95	140935	48.45	ug/l	0.00
Spiked Amount			Recovery	=	96.90%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	3.04	59	18416	35.639	ug/l	95
31) Cyclohexane	5.59	56	15964	5.299	ug/l	99
39) Methylcyclohexane	7.46	83	6754	1.908	ug/l	96
40) Benzene	6.14	78	702585	84.005	ug/l	99
42) 1,2-Dichloroethane	6.20	62	10105	3.475	ug/l	94
69) o-Xylene	10.69	106	44170	11.064	ug/l	100
73) Isopropylbenzene	11.02	105	3736	0.380	ug/l	96
95) Naphthalene	13.83	128	7406	0.794	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072619\
Data File : VX011085.D
Acq On : 26 Jul 2019 13:05
Operator : JC/SP
Sample : K4014-02 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2

Quant Time: Jul 27 02:34:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 24 05:39:23 2019
Response via : Initial Calibration

