

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072619\
 Data File : VX011090.D
 Acq On : 26 Jul 2019 15:02
 Operator : JC/SP
 Sample : K4014-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-5

Quant Time: Jul 27 03:23:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 24 05:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	202234	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	321599	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	296685	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	138936	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	110014	51.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.26%	
35) Dibromofluoromethane	5.50	113	100520	49.24	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.48%	
50) Toluene-d8	8.71	98	384567	50.23	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.46%	
62) 4-Bromofluorobenzene	11.13	95	135841	48.27	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.54%	
Target Compounds						
16) Acetone	2.45	43	2036	2.095	ug/l #	Qvalue 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072619\
Data File : VX011090.D
Acq On : 26 Jul 2019 15:02
Operator : JC/SP
Sample : K4014-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-5

Quant Time: Jul 27 03:23:43 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 24 05:39:23 2019
Response via : Initial Calibration

