

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072619\
 Data File : VX011093.D
 Acq On : 26 Jul 2019 16:12
 Operator : JC/SP
 Sample : K4017-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SU19-MSW0106

Quant Time: Jul 27 03:37:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 24 05:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	208490	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	327957	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	306921	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	145212	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	115103	51.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.78%	
35) Dibromofluoromethane	5.50	113	103215	49.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.16%	
50) Toluene-d8	8.71	98	400003	51.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.46%	
62) 4-Bromofluorobenzene	11.13	95	139001	48.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.88%	
Target Compounds						
					Qvalue	
16) Acetone	2.44	43	2899	2.894	ug/l #	79
31) Cyclohexane	5.59	56	1955	0.649	ug/l	86
39) Methylcyclohexane	7.46	83	1266	0.362	ug/l	92
67) Ethyl Benzene	10.25	91	3390	0.316	ug/l	94
68) m/p-Xylenes	10.35	106	2550	0.620	ug/l	95
84) 1,2,4-Trimethylbenzene	11.80	105	3140	0.380	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072619\
Data File : VX011093.D
Acq On : 26 Jul 2019 16:12
Operator : JC/SP
Sample : K4017-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SU19-MSW0106

Quant Time: Jul 27 03:37:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 24 05:39:23 2019
Response via : Initial Calibration

