

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072619\
 Data File : VX011095.D
 Acq On : 26 Jul 2019 16:59
 Operator : JC/SP
 Sample : PB121723ZHE#01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB121723ZHE#01

Quant Time: Jul 27 03:40:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 24 05:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	205252	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	324689	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	317243	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	145830	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	110381	50.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.10%	
35) Dibromofluoromethane	5.50	113	101335	49.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.34%	
50) Toluene-d8	8.71	98	403338	52.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.36%	
62) 4-Bromofluorobenzene	11.13	95	143390	50.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.94%	
Target Compounds						
16) Acetone	2.44	43	12630	12.807	ug/l	96
20) Methylene Chloride	2.85	84	2599	1.244	ug/l	93
43) Isopropyl Acetate	6.45	43	12080	2.690	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072619\
Data File : VX011095.D
Acq On : 26 Jul 2019 16:59
Operator : JC/SP
Sample : PB121723ZHE#01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB121723ZHE#01

Quant Time: Jul 27 03:40:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 24 05:39:23 2019
Response via : Initial Calibration

