

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072619\
 Data File : VX011096.D
 Acq On : 26 Jul 2019 17:22
 Operator : JC/SP
 Sample : PB121723ZHE#02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB121723ZHE#02

Quant Time: Jul 27 03:43:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 24 05:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	203865	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	328034	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	319051	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	150095	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	112167	51.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.44%	
35) Dibromofluoromethane	5.50	113	102876	49.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.82%	
50) Toluene-d8	8.71	98	403658	51.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.38%	
62) 4-Bromofluorobenzene	11.13	95	146622	51.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.16%	
Target Compounds						
16) Acetone	2.44	43	12684	12.949	ug/l	91
20) Methylene Chloride	2.86	84	2792	1.345	ug/l #	93
43) Isopropyl Acetate	6.45	43	11557	2.547	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072619\
Data File : VX011096.D
Acq On : 26 Jul 2019 17:22
Operator : JC/SP
Sample : PB121723ZHE#02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB121723ZHE#02

Quant Time: Jul 27 03:43:04 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 24 05:39:23 2019
Response via : Initial Calibration

