

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072619\
 Data File : VX011099.D
 Acq On : 26 Jul 2019 18:32
 Operator : JC/SP
 Sample : PB121723ZHE#05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB121723ZHE#05

Quant Time: Jul 27 03:48:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 24 05:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	194073	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	309181	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	297429	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	136626	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	105200	50.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.90%	
35) Dibromofluoromethane	5.50	113	96597	49.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.44%	
50) Toluene-d8	8.71	98	380288	51.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.34%	
62) 4-Bromofluorobenzene	11.13	95	133418	49.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.62%	
Target Compounds						
16) Acetone	2.45	43	12513	13.419	ug/l	95
20) Methylene Chloride	2.86	84	2727	1.380	ug/l #	83
43) Isopropyl Acetate	6.45	43	11540	2.699	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072619\
Data File : VX011099.D
Acq On : 26 Jul 2019 18:32
Operator : JC/SP
Sample : PB121723ZHE#05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB121723ZHE#05

Quant Time: Jul 27 03:48:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 24 05:39:23 2019
Response via : Initial Calibration

