

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072619\
 Data File : VX011102.D
 Acq On : 26 Jul 2019 19:43
 Operator : JC/SP
 Sample : PB121723ZHE#08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB121723ZHE#08

Quant Time: Jul 27 03:59:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 24 05:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	201385	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	322850	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	313197	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	144383	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	110148	51.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.82%	
35) Dibromofluoromethane	5.50	113	101401	49.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.96%	
50) Toluene-d8	8.71	98	402471	52.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.74%	
62) 4-Bromofluorobenzene	11.13	95	142226	50.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.68%	
Target Compounds						
16) Acetone	2.44	43	468818	484.504	ug/l	98
20) Methylene Chloride	2.86	84	152930	74.578	ug/l	96
43) Isopropyl Acetate	6.45	43	12321	2.759	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072619\
Data File : VX011102.D
Acq On : 26 Jul 2019 19:43
Operator : JC/SP
Sample : PB121723ZHE#08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB121723ZHE#08

Quant Time: Jul 27 03:59:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 24 05:39:23 2019
Response via : Initial Calibration

