

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072619\
 Data File : VX011105.D
 Acq On : 26 Jul 2019 20:53
 Operator : JC/SP
 Sample : PB121723ZHE#11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB121723ZHE#11

Quant Time: Jul 27 04:06:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 24 05:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	192911	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	311115	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	297320	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	132910	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	107100	52.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.38%	
35) Dibromofluoromethane	5.50	113	97473	49.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.72%	
50) Toluene-d8	8.71	98	386510	52.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.38%	
62) 4-Bromofluorobenzene	11.13	95	133589	49.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.14%	
Target Compounds						
16) Acetone	2.44	43	453804	489.589	ug/l	96
20) Methylene Chloride	2.86	84	148209	75.451	ug/l	97
43) Isopropyl Acetate	6.45	43	11624	2.702	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072619\
Data File : VX011105.D
Acq On : 26 Jul 2019 20:53
Operator : JC/SP
Sample : PB121723ZHE#11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB121723ZHE#11

Quant Time: Jul 27 04:06:41 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 24 05:39:23 2019
Response via : Initial Calibration

