

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX072619\
 Data File : VX011114.D
 Acq On : 27 Jul 2019 01:33
 Operator : JC/SP
 Sample : K4017-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SU19-MSW0107

Quant Time: Jul 27 04:38:25 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 24 05:39:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	192191	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	308826	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	290124	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	133975	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	106101	51.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.78%	
35) Dibromofluoromethane	5.49	113	96104	49.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.06%	
50) Toluene-d8	8.71	98	368404	50.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.22%	
62) 4-Bromofluorobenzene	11.13	95	130881	48.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.86%	
Target Compounds						
16) Acetone	2.45	43	2802	3.034	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX072619\
Data File : VX011114.D
Acq On : 27 Jul 2019 01:33
Operator : JC/SP
Sample : K4017-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SU19-MSW0107

Quant Time: Jul 27 04:38:25 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X071619W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 24 05:39:23 2019
Response via : Initial Calibration

